


Fluid transfer

Quick connector



CUSTOMER APPROVAL READ & APPROVED:

Name: _____

Date: __/__/__

Signature: _____

Info on Packaging & Gamma Irradiation

Packaging: double bagged
Label: on outside bag
Gamma Irradiation: Yes, 25-40kGy

Bill of Materials

All connections, unless otherwise specified, are reinforced with single cable ties

This document is the property of Hitma UltraPure B.V. and must not be disclosed to third party without the written consent of Hitma UltraPure B.V. .This drawing represents the components only schematically to illustrate the function and/or assembly of the equipment/device and to enable the customer to check the component's completeness. The semblance of the real components may therefore deviate from the drawing.